

Method comparison: H. pylori: HM-CAP EIA, H. pylori: Immunochromatic

CLSI EP12-A2 - Example 10.3.2

Last updated 23 June 2015 at 15:05 by Simon Huntington

Frequencies

N		536		
H. pylori: HM-CAP EIA	H. pylori: Immunochromatic		Total	
	-	+		
-	222	15	237	
	0.414	0.028	0.442	
+	14	285	299	
	0.026	0.532	0.558	
Total	236	300	536	
	0.440	0.560	1.000	

Proportions

Agreement	Proportion	Wilson 95% CI
-	0.937	0.898 to 0.961
+	0.953	0.923 to 0.972
Total	0.946	0.923 to 0.962